Zhermack

SILICONES - CERTIFICATE OF ANALYSIS

B220005 ROYAL BITE 2x50ML

MANUFACTURED DATE: 2025-04-28	BATCH NUMBER: 0000469417	EXPIRY DATE: 2028-04-26

Test parameter	Test Result	Specification	Test method
Room temperature (°C)	22,8	[22,0-24,0]	2100-WI QC-001
Relative Humidity (%RH)	47,9	[40,0-60,0]	2100-WI QC-001
Base temperature (°C)	23,0	[22,0-24,0]	2100-AM QC-023
Catalyst temperature (°C)	23,1	[22,0-24,0]	2100-AM QC-023
Mixing time (s)	Compliant	Automatic	

Product Characteristic	Test Result	Specification	Test method
Colour base	Compliant	Coral	2100-AM QC-004
Colour catalyst	Compliant	White	2100-AM QC-004
Visual check	Compliant	Homogeneus*	2100-AM QC-004
Mixed material	Compliant	Homogeneus*	2100-AM QC-004
Setting time 35°C (s)	252	160 - 280	2100-AM QC-023
Hardness 8' (23°C) (ShA)	85	83 - 93	2100-AM QC-013
Hardness 1h (23°C) (ShA)	86	85 - 95	2100-AM QC-013

Visual Inspection Report

Finished product inspection result: COMPLIANT.

Badia Polesine, 2025-05-09 QUALITY CONTROL DEPARTMENT

*Homogeneus and free of extraneous matter

This is an electronic document, the signature is not necessary.

Zhermack S.p.A.

Via Bovazecchino, 100
45021 Badia Polesine (RO) Italy
T +39 0425 597611 | F +39 0425 53596
www.zhermack.com | info@zhermack.com

C.F. e P.IVA IT00594630295
R.E.A. 86603 - R.I. RO 00594630295
M. RO 001381 - Isc. Reg. A.E.E. IT07111000002
Cap. Soc. € 1.032.000,00